


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVX-FD1-PS.

S.N.: 8302

Date: Feb 3/98



- a) Output signal amplitude:  
TTL,  $R_L \geq 50\Omega$
- b) Pulse width:  
50ns to 50 $\mu$ s
- c) Division Factor: 2 to 999
- d)
- e) PRF: 0 to 10 MHz
- f) Jitter, stability: OK
- g) Prime power: 120/240V  
50-60Hz

Top: Output,  $R_L = 50\Omega$ ,  $\pm 10$

Bot: Input, 10 MHz