


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMH-1-P*

S.N.: *9303*

Date: *DEC 1 1997*



- a) Output signal amplitude: *0 TO +10V*
- b) Pulse width: *≤ 200ns*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+24V DC*

40dB ATTEN
10 VOLTS/DIV
PRF = 5.0 MHz

+24V DC
400 mA MAX