


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMH-1-P*

S.N.: *8304*

Date: *DEC 1 1997*



40 db. ATTEN
10 VOLTS/DIV
PRF = 5 MHz

- a) Output signal amplitude:
0 TO +18V
- b) Pulse width:
≤ 200 pS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 VDC
400 mA (MAX)