


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV-151B-C
S.N.: 8306
Date: DEC 15 1997



- a) Output signal amplitude: 0 TO 200 V
- b) Pulse width: $R_L > 50 \Omega$
SINE, SQUARE, TRIANGLED
- c) Rise time: 10% OFFSET
 $\leq 1 \mu s$
- d) Fall time:
 $\leq 1 \mu s$
- e) PRF: 0 TO 100 KHz
- f) Jitter, stability: OK
- g) Prime power: 170/240 V
50-60 Hz

10 μs / DIV
50 V / DIV
 $R_L = 100 \Omega$