


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PA-151B-C*
S.N.: *8307*
Date: *DEC 15 1997*



- a) Output signal amplitude: *0 TO ± 200 V*
- b) Pulse width: *($R_L > 50 \Omega$)
SINE, SQUARE, TRIANGLE*
- c) Rise time: *+DC OFFSET
≤ 1 μ S*
- d) Fall time: *≤ 1 μ S*
- e) PRF: *0 TO 100 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V
50-60 Hz*

*10 μ S/DIV
50 V/DIV
 $R_L = 50 K$*