


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-24-C-P*

S.N.: *8308*

Date: *DEC 11 1997*



- a) Output signal amplitude: *0 TO +100 V*
- b) Pulse width: *1 TO 5 ns*
- c) Rise time: *≤ 0.4 ns*
- d) Fall time: *≤ 0.4 ns*
- e) PRF: *0 TO 100 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *100V ± 10%*

60 dB ATTEN
100 V/5ns/10W
PRF = 100 kHz