


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-S-HV1-C-P*

S.N.: *8312*

Date: *JAN 6 1998*



- a) Output signal amplitude: *0 TO +40 V*
- b) Pulse width: *≤ 350 ps (FWHM)*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz

50 dB ATTEN

32 VOLTS/DIV

PRF = 100 kHz