


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVAMP-3-P-C-OT-01*
 S.N.: *8314*
 Date: *MAR 4 1998*



- a) Output signal amplitude: *0 TO +20 VOLTS*
- b) Pulse width: *(50 ns)*
- c) Rise time: *10 NS TO 100 NS*
≤ 200 PS
- d) Fall time: *≤ 300 PS*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 VA
- h) offset: *0 TO ±5 V*

40 db ATTEN
10 VOLTS/DIV
PRF = 100 kHz
TRIG 5 ns/DIV
150 I 500 PS/DIV
(RISE TIME)

[Signature]