


PULSE GENERATOR
PERFORMANCE CHECK

Model: AP0-7D-C-ET-AT-PN-05-M2-0P1B
-NW4A

S.N.: 8319


Date: MAY 21 1998



A
Fast TOP MONITOR OUT TO 10K
1V/DIV

BOI VLAND, 10V/DIV PRF=
RLOM = 0.55 20Hz

- a) Output signal amplitude:
0 TO ± 100 AMP
- b) Pulse width:
(TO 0 TO ± 100 μ)
2 TO 200 μS
- c) Rise time:
(0.1% MAX DUTY CYCL)
≤ 14 ns
- d) Fall time:
≤ 14 ns
- e) PRF: 0 TO 200 Hz
(0.1% MAX DUTY CYCL)
- f) Jitter, stability: 0%



B
NonT AS ⊕ BUT NonT.

- g) Prime power: 120/240 V
50-60 Hz
(0.9 AMP)

[Handwritten signature]