


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MR-61-PN-05*



S.N.: *8335*

Date: *FEB 26 1998*



- a) Output signal amplitude: *0 TO ± 250 V*
- b) Pulse width: *($R_L \geq 10\Omega$)
100 NS TO 1 MS
(20% MAX DUTY CYCLE)*
- c) Rise time: *≤ 10 NS*
- d) Fall time: *≤ 10 NS*
- e) PRF: *0 TO 1 KHz
(20% MAX DUTY)*
- f) Jitter, stability: *Cycle*
- g) Prime power: *OK
120/240 V
50-60 Hz*

*$R_L = 10K$
50 V/DIV*



MA