


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ARR-A-2-PW-PS-P*

S.N.: *8336*

Date: *MAR 2 1998*



- a) Output signal amplitude: *0 TO +50 VOLTS*
- b) Pulse width: *20 TO 200 NS*
- c) Rise time: *(20% MAX DUTY CYCLE) ≤ 5 NS*
- d) Fall time: *≤ 5 NS*
- e) PRF: *0 TO 10 MHz. (20% MAX DUTY CYCLE)*
- f) Jitter, stability: *OK*
- g) Prime power: *100/200V*
50-60 Hz.

50 db ATTN
232 VOLTS/DIV
PRF ≈ 1.0 MHz

[Signature]