


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MV-1010-C*


S.N.: *8342*

Date: *MAR 13 1987*




- a) Output signal amplitude:  
*0 TO ± 100 V*
- b) Pulse width: *( $R_L \geq 50 \Omega$ )*  
*50 NS TO 10 MS*
- c) Rise time: *(10% MAX DUTY CYCLE)*  
*≤ 15 NS*

- d) Fall time:  
*≤ 15 NS*
- e) PRF: *0 TO 1.0 MHz*



- f) Jitter, stability:  
*OK*
- g) Prime power: *100/200 V 50-60*  
*± 10%*



*(B) Next 30 V KANLES*  
 *$R_L = 50 \Omega$ , 10V/DIV*  
*500 NS/DIV*

*[Signature]*