


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-6B-C-P*


S.N.: *8349*

Date: *MAR 6 1998*




- a) Output signal amplitude: *0 to 2 Amps*
- b) Pulse width: *50ns to 5μs*
- c) Rise time: *≤ 10ns*
- d) Fall time: *≤ 10ns*
- e) PRF: *0 to 10kHz*
- f) Jitter, stability: *OK*

Ⓐ *AVO-6B-T removed*
R_L = 50Ω, 20V/DM, 1μs/DM



- g) Prime power: *120/2000*
50-60Hz

[Signature]



Ⓑ *AVO-6B-T output*
1 V/DM
1 μs/DM