

PULSE GENERATOR  
PERFORMANCE CHECK

Model: AV-155C-C-P-AB1

S.N.: 8353

Date: March 23, 1998

Waveforms:


see attached printouts of

- #1) Channel A + monitor
- #2) Channel B + monitor
- #3) A, B and AB delay

- a) Output signal amplitude:  
0.00 - 5.00A
- b) Pulse width: 0.0 - 9.9 ms
- c) Rise time:  $< 10\mu s$
- d) Fall time:  $< 10\mu s$
- e) PRF: 0 - 1 Hz
- f) Jitter, stability: OK
- g) Prime power:  
120/240V, 50/60Hz
- h) AB Delay  
000.0 - 999.9 ms

Waveform #1

Tek Run: 2MS/s Sample **100ns**



Waveform #2

Tek Run: 20ks/s Sample **11198**

