


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-61-C-RPCIB1-N*

S.N.: *8364*


Date: *MAR 25 1998*



- a) Output signal amplitude: *0 TO -500 V (R_L > 1.2Ω)*
- b) Pulse width: *300 μs TO 3 ms*
- c) Rise time: *(50% MAX DUTY CYCLE)*
≤ 40 NS
- d) Fall time: *≤ 40 NS*
- e) PRF: *0 TO 1 KHz*
(50% MAX DUTY CYCLE)
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz

R_L = 1K

100 V/DIV



[Signature]