


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV6-2-P

S.N.: 8367

Date: MAR 18 1998



- a) Output signal amplitude: 0 TO +120 VOLTS
- b) Pulse width: $\leq 2 \text{ NS}$
- c) Rise time: $\leq 1 \text{ NS}$
- d) Fall time: $\leq 1 \text{ NS}$
- e) PRF: 0 TO 50 KHz
- f) Jitter, stability: OK
- g) Prime power: +15 VDC
350 mA
MAX.

60 db ATTEN
- 100 VOLTS/DIV
2 NS/DIV
PRF = 30 KHz.

Signature