


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AV-2-C-P-M*
S.N.: *8369*
Date: *MAY 12 1998*



- a) Output signal amplitude: *0 TO +10 VOLTS*
- b) Pulse width: *(50ns)*
- c) Rise time: *2 TO 50 NS*
≤ 100ps
- d) Fall time: *≤ 135ps*
- e) PRF: *0 TO 50 KHz*
- f) Jitter, stability: *O/L*
- g) Prime power: *120/240V*
50-60 Hz

40 db ATTN
10 VOLTS/DIV
TOP 5 NS/DIV
SOI 100 PS/DIV
(RISE TIME)
50 KHz PRF