


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVIMP-3-P-CED1*  
 S.N.: *3371*  
 Date: *Feb 9 1998*



- a) Output signal amplitude:  
*> +20 VOLTS*
- b) Pulse width: *50 ns*  
*87 ns*
- c) Rise time: *≤ 300 ps*
- d) Fall time: *≤ 300 ps*
- e) PRF: *0 to 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power: *+24 ± 4.0 VOLTS*  
*350 mA.*

*[Signature]*

*40 dB ATTEN*  
*10 VOLTS/DIV*  
*10 ns/DIV*  
*100 ps/DIV*  
*(RISE TIME)*  
*PRF = 300 kHz*