


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMP-3-6501*

S.N.: *8372*

Date: *FEB 9 1998*



- a) Output signal amplitude: *> +26 V_{0.45} TO*
- b) Pulse width: *50 ns*
- c) Rise time: *87 ns*
≤ 300 ps
- d) Fall time: *≤ 300 ps*
- e) PRF: *0 TO 10 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+24 ± 4.0 V_{0.45}*
350 mA