


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMP-3-P-6E01*

S.N.: *8373*

Date: *FEB 10 1998*



- a) Output signal amplitude:
> +20 VOLTS
- b) Pulse width: *(50%)*
87 NS
- c) Rise time:
≤ 300 ps
- d) Fall time:
≤ 300 ps
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability:
OK
- g) Prime power:
+24 ± 4.0 VOLTS
350 mA (MAX)

40 db ATTN
10 VOLTS/DIV
TR 10 NS/DIV
BOY 100 ps/DIV
(KLSB TIME)
PRF ≈ 300 kHz