


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMP-3-P-GE01*

S.N.: *8374*

Date: *FEB 10 1998*



- a) Output signal amplitude:
± 20 VOLT
- b) Pulse width: *(50%)*
87 NS
- c) Rise time:
≤ 300 PS
- d) Fall time:
≤ 300 PS
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*

- g) Prime power:
+24 ± 4.0 VOLT
350 mA (RMS)

40 dB ATTEN
10 VOLT/DIV
TOP 10 NS/DIV
BOT 100 PS/DIV
(RISE TIME)
PRF ≈ 300 kHz

[Signature]