


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AN-ET-C-PN-DA-EW*

S.N.: *8385*

Date: *MAY 29 1998*



- a) Output signal amplitude: *0 TO ±100 VOLTS*
- b) Pulse width: *(50N)*
- c) Rise time: *1 TO 5 NS*
≤ 0.4 NS
- d) Fall time: *≤ 0.4 NS*
- e) PRF: *0 TO 100K Hz*
- f) Jitter, stability: *61L*
- g) Prime power: *120 / 240 V*
50-60 Hz

60 db ATTEN

100 VOLTS/DIV

PRF = 20K Hz.

[Signature]