


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVI-U-1A3-CP

S.N.: 8386

Date: FEB 23 1998



- a) Output signal amplitude: 0 TO +100 VOLTS
- b) Pulse width: 150ns
- c) Rise time: 4 TO 100 NS  
≤ 200 NS
- d) Fall time: ≤ 4 NS
- e) PRF: 0 TO 5 KHz
- f) Jitter, stability: OK
- g) Prime power: 100 ± 10 %  
50-60 Hz

60 db ATTEN  
- - 100 VOLTS/DIV  
TOP 5 NS/DIV  
BOT 500 PS/DIV  
(RISE TIME)