


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AM-4B-PW-PS-PN-AT-PWT*


S.N.: *8389*

Date: *MAR 23 1998*



- a) Output signal amplitude:
0 TO ± 400 V (RL=7.50 Ω)
- b) Pulse width: *0.1 TO 100 NS*
0.5% DUTY CYCLE
- c) Rise time:
≤ 20 NS
- d) Fall time:
≤ 20 NS
- e) PRF: *0 TO 10 KHz*
0.5% DUTY CYCLE
- f) Jitter, stability:
OK
- g) Prime power: *120 / 240 V*
50 - 60 Hz

Ⓐ *Power, RL = 50 Ω*
20 ns/div
150 V/div



Ⓑ *Narr RL = 50 Ω*
100 NS/div, 15.0V/div

[Handwritten signature]