




PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-1011-C*
 S.N.: *8397*
 Date: *MAR 17 1998*



- a) Output signal amplitude: *0 TO ±100 VOLTS*
- b) Pulse width: *($R_L \ge 50\Omega$)
100NS TO 1 MS*
- c) Rise time: *(10% MAX DUTY CYCLE)
≤ 10 NS*
- d) Fall time: *610NS*
- e) PRF: *0 TO 1.0 MHz
(10% MAX DUTY CYCLE)*
- f) Jitter, stability: *OK*
- g) Prime power: *120/1240V
50-60 MHz*



Ⓑ *Nout* *R_{out} = 50 Ω , R_L = 50 Ω*
10V/DIV, 100ns/DIV
PRF = 1.0 kHz