


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-N-1-C-P-05*

S.N.: *8401*

Date: *MAY 11 1968*



- a) Output signal amplitude: *0 TO +10 VOLTS*
- b) Pulse width: *0.2 TO 4.0 NS*
- c) Rise time: *≤ 100 PS*
- d) Fall time: *≤ 135 PS*
- e) PRF: *0 TO 100 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/280V
50-60Hz*

40 db ATTEN
10 VOLTS/DIV
TR 1 NS/DIV
BDI 100 PS/DIV
(RISE TIME)
PRF = 100 kHz