


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMP-1A-C-M-D*

S.N.: *8404*

Date: *APRIL 21 1968*



- a) Output signal amplitude: *0 to +5 Volts*
- b) Pulse width: *5 to 100 NS*
- c) Rise time: *≤ 100 PS*
- d) Fall time: *≤ 135 PS*
- e) PRF: *0 to 1 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*

*50-60 Hz*  
*4/8*

*40 dB ATTN*  
*5 Volts/Div*  
*Top 5 ns/Div*  
*Bot 100 ps/Div*  
*(RISE TIME)*  
*PRF = 1.0 MHz*