


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-MU-C-P*

S.N.: *8407*

Date: *MAY 25 1998*



- a) Output signal amplitude:
0 TO +15 V (S.O.W)
- b) Pulse width:
2 TO 50 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability: *OK*
- g) Prime power:
120/240 V
50-60 Hz

40db ATTN
10 VOLTS/DIV
TOP: 5 NS/DIV
BOT: 200 PS/DIV.
PRF = 30 KHz.