


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2-C-N*

S.N.: *8412*

Date: *MAY 29 1998*



- a) Output signal amplitude:
0 to 350 V
- b) Pulse width: *(50 ns)*
5 to 100 ns
- c) Rise time:
≤ 2 ns
- d) Fall time:
≤ 3 ns
- e) PRF: *0 to 5 kHz*
- f) Jitter, stability: *OK*

70 db ATTEN
= 320 VOLTS/DIV
PRF = 2 kHz

g) Prime power: *100 ± 10%*
VOLTS
50-60 A