


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVD-8C-CN*

S.N.: *8424*

Date: *JUNE 3 1998*



- a) Output signal amplitude:
0 to 40V (to $R_L > 0.2\Omega$)
- b) Pulse width:
200 Amplitude limit
2 μ S to 20 ms
- c) Rise time: *(\neq DC)*
 $\leq 1 \mu$ S
- d) Fall time:
 $\leq 1 \mu$ S
- e) PRF:
0 to 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:
 - a) *120/240 V*
 - 50-60 Hz*
 - DC LAB PS*
 - 0 to 40*
 - 100 AmPS.*

$R_L = 0.2 \Omega$
PMF = 100 kHz