


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AV6-3A-C091

S.N.: 8437

Date: July 8 1998



- a) Output signal amplitude:  
+150 TO +450V
- b) Pulse width:  
7 NS
- c) Rise time:  
≤ 3NS
- d) Fall time:  
≤ 3NS
- e) PRF:  
0 TO 10 K Hz
- f) Jitter, stability:  
OK
- g) Prime power:

+ 15 VOLTS  
(400 mA)  
max

60 dB ATTEN:  
100 VOLTS/DIV