


PULSE GENERATOR
PERFORMANCE CHECK


Model: AM-7B-C-0P1B-PN

S.N.: 8438

Date: JUNE 25 1994



A) 150V/DIV, 100ns/DIV
Pulse, $R_L = 50\Omega$, PRF = 2 kHz



B) 150V/DIV, 10us/DIV
Nons. $R_L = 50\Omega$, PRF = 50 Hz

- a) Output signal amplitude:
0 TO ± 700 VOLTS
- b) Pulse width: ($R_L \geq 50\Omega$)
100ns TO 100us
- c) Rise time: 0.5% MAX DUTY CYCLE
 ≤ 50 NS
- d) Fall time: ≤ 50 NS
- e) PRF: 0 TO 10 KHz
(0.5% MAX DUTY CYCLE)
- f) Jitter, stability: OK
- g) Prime power: 120/240V
50-60 Hz

[Handwritten signature]

