


PULSE GENERATOR
PERFORMANCE CHECK



Model: *MV-1010-C*

S.N.: *8441*

Date: *June 27 1998*



- a) Output signal amplitude: *0 TO ±100 VOLTS ✓*
- b) Pulse width: *(R_L = 50Ω)
50 NS TO 10 MS*
- c) Rise time: *(10% MAX DUTY CYCLE)
≤ 15 NS ✓*
- d) Fall time: *≤ 15 NS ✓*
- e) PRF: *0 TO 1.0 MHz ✓
(10% MAX DUTY CYCLE)*
- f) Jitter, stability: *OK ✓*
- g) Prime power: *120/240 V
50 to 1 Hz*



Ⓑ *NEXT 30 V RANGE
R_L = 50Ω 10V/DIV
500 NS/DIV*

[Signature]