


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-SA-C-P

S.N.: 8443

Date: Jul 15 1998



- a) Output signal amplitude:
0 TO +20 VOLT
- b) Pulse width:
 $\leq 200ps$
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: 0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

40 db ATTN
10 VOLT/DIV
PRF \approx 100 KHz