


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVH-5-1-C-P*

S.N.: *8444*

Date: *JULY 28 1998*



- a) Output signal amplitude: *0 TO +10 VOLTS*
- b) Pulse width: *≤ 130 ps*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*

40 dB ATTEN
10 VOLTS/DIV
20V 1NS/DIV
200ps/DIV
PRF ≈ 1.0 MHz

516012
[Signature]