


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVIR-1-PS-N-AT-PWT*

S.N.: *8520*

Date: *AUG 20 1998*



- a) Output signal amplitude: *0 to -200 volts*
- b) Pulse width: *($R_L = 50\Omega$)*
- c) Rise time: *1 to 5 ns*
 $\leq 0.5 ns$
- d) Fall time: *$\leq 2 ns$*
- e) PRF: *0 to 50 kHz*
- f) Jitter, stability: *OK*

g) Prime power: *120/240V*
50-60Hz
[Signature]

60 dB ATTEN
100 volts/div
TOP 2 ns/div
PRF = 30 kHz