


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-5-1-C-N

S.N.: 8535

Date: Aug 17 1998



- a) Output signal amplitude: 0 to -10 V_{0.7A3}
- b) Pulse width: ≤ 130 ps
- c) Rise time: N/A
- d) Fall time: N/A
- e) PRF: 0 to 1.0 MHz
- f) Jitter, stability: OK
- g) Prime power: 120/240 V
50-60 Hz

40 db ATTN
10 V_{0.7A3} / DIV
TOP 1.0 ns / DIV
BOT 200 ps / DIV
PRF = 100 kHz