


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-V-143-P

S.N.: 8541

Date: JUN 7 10 1998



- a) Output signal amplitude:
0 TO +100V (70.50V)
- b) Pulse width:
5 TO 100NS
- c) Rise time:
≤ 200ps
- d) Fall time:
≤ 4NS
- e) PRF:
0 TO 5KHz
- f) Jitter, stability:
OK

60 db ATTEN
100 VOLTS/DIV
20P 10 NS/DIV
MID 1 NS/DIV
BOT 200 PS/DIV } RISE TIME

g) Prime power:
+15V DC
400 mA (MAX)

[Signature]

PRF = 5 kHz