


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2-B-PN-05*

S.N.: *8551*

Date: *AUG 31 1998*



60 db ATTN

100 VOLTS/DIV

TOP P OUT

BOT N OUT

PRF = 2.0 KHz

2 10 NS/DIV

- a) Output signal amplitude:
0 TO ± 350 V ($R_L = 50\Omega$)
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 4 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz