


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVR4-1-C-7-AT*

S.N.: *8553*

Date: *AUG 28 1998*



- a) Output signal amplitude: *0 TO 1100V ( $R_L > 200\Omega$ )*
- b) Pulse width: *0.25 TO 50  $\mu$ S*
- c) Rise time:  *$\leq 50$  NS*
- d) Fall time:  *$\leq 50$  NS*
- e) PRF: *0 TO 1 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*  
*50-60Hz*

*500V/DIV*  
 *$R_L = 10K$*

