


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-A-1-PW-C-PN-W1-AT-PWT

S.N.: 8563

Date: Jul 4 1998



- a) Output signal amplitude: 0 TO ± 200 V (200V)
- b) Pulse width: $R_L \geq 50 \Omega$
50 NS TO 1.0 US
(0.5% MAX DUTY CYCLE)
- c) Rise time: ≤ 5 NS
- d) Fall time: ≤ 5 NS
- e) PRF: 0 TO 100 K Hz
(0.5% MAX DUTY CYCLE)
- f) Jitter, stability: OK
- g) Prime power: 120/240 V
50-60 Hz

$R_L = 50 \Omega$

PRF = 10 K Hz