


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AMR-4A-PW-C-PN-0P1B*

S.N.: *8565*

Date: *Oct 27 1998*



Ⓐ *Test  $R_L = 50 \Omega$*



Ⓑ *Next  $R_L = 50 \Omega$*

- a) Output signal amplitude: *0 TO 1400 V ( $R_L = 50 \Omega$ )*
- b) Pulse width: *50 NS TO 5.0 US*
- c) Rise time: *0.5% MAX DUTY CYCLE*  
 *$\leq 20 NS$*
- d) Fall time:  *$\leq 20 NS$*
- e) PRF: *0 TO 10 K Hz*  
*(0.5% MAX DUTY CYCLE)*
- f) Jitter, stability: *OK*
- g) Prime power: *150/240 V*  
*50-60 Hz*

*[Signature]*