


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PU-110B-PS-BTLB*

S.N.: *8569*

Date: *SEPT 10 1998*



*NON INVERTING*

*INPUT  
4 Vpp*

*OUTPUT  
200 Vpp  
WITH +50V  
OFFSET.*


- a) Output signal amplitude:  
*0 TO ± 200 Vpp*
- b) GAIN *(R<sub>L</sub> ≥ 50K)*  
*+x1 TO +x100 AND*  
*-x1 TO -x100*
- c) Rise time:

*≤ 1 µs.*

e) PRF: *DC TO 100K Hz*

f) Jitter, stability: *OK*

g) Prime power: *120/240V*  
*50-60 Hz*



*INVERTING.*

*INPUT  
4 Vpp*

*OUTPUT  
200 Vpp  
WITH +50V  
OFFSET.*

*R<sub>L</sub> = 100K.*