


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ARR-95-C*

S.N.: *8571*

Date: *SEPT 22 1998*



- a) Output signal amplitude:
0 TO 440 VOLTS (P TO P)
- b) Pulse width: *R_L ≥ 100K*
50% DUTY CYCLE.

c) Rise time:
≤ 20 NS

d) Fall time:
≤ 20 NS

e) PRF: *1 KHz TO 100 KHz.*

f) Jitter, stability:
OK

g) Prime power:
120/240V
50 to 60 Hz.

2 ns/DIV
50 V/DIV

2 ns/DIV
50 V/DIV

SCOPE PROBE
NOT PROPERLY
COMPENSATED.