


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVIR-1-P5-PN-05-ECL-1FO1CA*

S.N.: *8583*

Date: *NOV 11 1998*



- a) Output signal amplitude:
0 to 200 V to 50mV
- b) Pulse width:
1 to 5 ns
- c) Rise time:
≤ 0.5 ns
- d) Fall time:
≤ 1.5 ns
- e) PRF:
0 to 50 kHz.
- f) Jitter, stability:
OK
- g) Prime power:

120/240V
50-60 Hz.
AK

60 dB ATTEN

-100 VOLTS/DIV

TOP P_{out}

BOT N_{out}

PRF = 10 kHz