


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-A7-S1-C*

S.N.: *8585*

Date: *OCT 15 1998*



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width: *($C_L = 50pF$)*
50NS TO 500NS
- c) Rise time: *(0.59. MAX DUTY CYCLE)*
 $\leq 2 NS$
- d) Fall time:
 $\leq 2 NS$
- e) PRF: *0 TO 100 KHz*
(0.59. MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:
100 \pm 10 VOLTS

60 db ATTN
100 VOLTS/DIV

50-60 KHz

