


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-C-N

S.N.: 8600

Date: OCT 20 1998



- a) Output signal amplitude: 0 to -30V (50Ω)
- b) Pulse width: ≤ 1 NS
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 to 1.0 MHz
- f) Jitter, stability: OK
- g) Prime power: 120/240V
50-60 Hz.

40 db ATTN
10 VOLT/PIV
PRF = 300 kHz