


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-M1-P-MA

S.N.: 2604

Date: OCT 1 1998



- a) Output signal amplitude: $+30 \text{ VOLT (50\Omega)}$
- b) Pulse width: 500 ns (FWHM)
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: $0.90 / \text{MTR}$
- f) Jitter, stability: OK
- g) Prime power: $+135 \pm 5 \text{ VOLT}$
 5 mA NVA

40 db ATTEN

10 VOLT/DIV

1.0 MTR PRF