


PULSE GENERATOR
PERFORMANCE CHECK

Model: ANMP-3-N-B

S.N.: 3608

Date: NOV 19 1998



- a) Output signal amplitude: 0 TO -20 V
- b) Pulse width: (50 ns)
8 TO 100 ns
- c) Rise time: ≤ 200 ps
- d) Fall time: ≤ 300 ps
- e) PRF: 0 TO 1.0 MHz
- f) Jitter, stability: C10 9, MAX DUTY CYCLE
OK
- g) Prime power: 120/240 V
5 TO 60 MHz

40 db ATTN.: 10 V/DIV
TOP 10 ns/DIV

MRP 1 ns/DIV (RISE TIME)

BDI 1 ns/DIV (FALL TIME)

