


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANN-4-C-P-LTA*

S.N.: *8635*

Date: *SEPT 15 1948*



- a) Output signal amplitude:
50-100 MHz : +20 VOLTS
150 MHz : +15 VOLTS
250 MHz : +5 VOLTS
- b) Pulse width:
0.2 TO 1.0 NS
- c) Rise time: *≤ 150 ps*
- d) Fall time:
≤ 150 ps
- e) PRF: *50 TO 250 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz

40 db ATTN
10 VOLTS / DIV
TOP 2 NS / DIV
BOT 20 ps / DIV

[Signature]